



SURAT PERINTAH KERJA

INSTALASI PEMELIHARAAN SARANA PRASARANA

NOMOR :

870/4/26/L/IPSR

KEPADA :


- | | | |
|----------------|---------|------------|
| 1. Pak Khoirul | 3. Dika | 5. Jesisca |
| 2. Pak Syaiful | 4. Aris | |

Selesai IPSRS

TANGGAL LAPORAN	NOMOR LAPORAN	PELAKSANAAN	LOKASI
2026-04-10 11:55:25	92/4/26	Minggu, 12 April 2026	Neonatus

NAMA ALAT	MERK	MODEL	NOMOR SERI
infant warmer : shvabe	-	-	-

LAPORAN	TINDAKAN	KETERANGAN
- Lampu infant warmer mati 1 buah	penggantian lampu infant warmer	penggantian lampu infant warmer



SPAREPART

- Lampu Halogen 220V Down 1
-
-

-
-
-
-
-
-

Mengetahui,

Teknisi Pelaksana

Pak Khoirul

Pak Syaiful

Dika

Aris

Jesisca

Penerima Pekerjaan

Kunti

Admin

Alivia Ayu Ninditya

bag. IPSRS RSUD SLG
Ka. IPSRS

Nova Dwi Hapsari
NIP. 198403012011012016