



SURAT PERINTAH KERJA

INSTALASI PEMELIHARAAN SARANA PRASARANA

NOMOR :

890/4/26/E/IPRSRS

KEPADA :


- | | | |
|----------------|---------|------------|
| 1. Pak Khoirul | 3. Aris | 5. Jesisca |
| 2. Pak Syaiful | 4. Dika | |

Selesai IPSRS

TANGGAL LAPORAN	NOMOR LAPORAN	PELAKSANAAN	LOKASI
2026-04-13 11:33:56	108/4/26	Rabu, 15 April 2026	Klinik Jantung

NAMA ALAT	MERK	MODEL	NOMOR SERI
Echocardiography	-	-	-

LAPORAN	TINDAKAN	KETERANGAN
echo tidak bisa ngeprint (minta segera di cek)	restart unit echo	



SPAREPART

- | | |
|----|----|
| 1. | 4. |
| 2. | 5. |
| 3. | 6. |

Mengetahui,

Teknisi Pelaksana

Bu Nova

Aris

Dika

Jesisca

Penerima Pekerjaan

Evi

bag. IPSRS RSUD SLG

Admin

Alivia Ayu Ninditya

Ka. IPSRS

Nova Dwi Hapsari
NIP. 198403012011012016